



Add value.
Inspire trust.

TÜV SÜD Product Service GmbH · Ridlerstr. 65 · 80339 Munich · Germany

Shenzhen Microprofit Biotech Co., Ltd
Room 1001 of Unit 2, Room 1001 and Room 1101 of Unit 1,
Building 2, Hongchuang Technology Center,
Xikeng Community, Fucheng Sub-district,
Longhua District, 518000, Shenzhen, Guangdong,
PEOPLE'S REPUBLIC OF CHINA

Your reference/letter of	Our reference/name	Tel. extension/Email	Fax extension	Date	Page
109172	713347729 GZ2443603_CL-S	medical_devices@tuvsud.com	---	2025-06-03	1 of 3

**TÜV SÜD Product Service GmbH
Confirmation Letter
CLI 109172 0005 Rev. 01**

Reference: 713347729 | GZ2443603_CL-S | GZ2443603_CL

To whom it may concern,

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2024/1860 amending Regulations (EU) 2017/746 (in the following referenced as IVDR) as regards the transitional provisions for certain in vitro diagnostic medical devices.

With this letter TÜV SÜD Product Service GmbH, designated under IVDR and identified by the number 0123 on NANDO, confirms that we have received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of IVDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of IVDR with the above stated manufacturer with the following Single Registration Number (SRN)

Single Registration Number: CN-MF-000019518

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below.

- Table 1 identifies the devices for which an IVDR application has been received, written agreement concluded and for which TÜV SÜD Product Service GmbH is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive.
- Table 2 identifies the devices for which an IVDR application has been received and a written agreement concluded, but TÜV SÜD Product Service GmbH has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive or these devices did not require a Notified Body certificate under Directives.

Registered Office: Munich
Trade Register Munich HRB 85742
UniCredit Bank GmbH · BIC HYVEDEMMXXX
IBAN DE13 7002 0270 0048 8522 11
VAT ID No. DE129484267
Information pursuant to § 2 [1] DL-InfoV
(Germany) at tuvsud.com/imprint

Supervisory Board:
Holger Lindner (Chairman)
Board of Management:
Walter Reithmaier (CEO)
Patrick van Welij

TÜV SÜD Product Service GmbH tuvsud.com/ps
Zertifizierstelle für Medizinprodukte / Certi- Hotline: +49 89 50084-747
fication Body for Medical Products
Ridlerstr. 65
80339 Munich
Germany





If devices covered by certificates issued under Directive Directive 98/79/EC (IVDD) that expired after 26. May 2022 and before 09. July 2024, without having been withdrawn, this letter also confirms that

- the manufacturer signed the written agreement under IVDR by the date of IVDD certificate expiry; or
- provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 54(1) of IVDR or Article 92(1) of the IVDR respectively.

The transition timelines in accordance Article 110 (3) of IVDR that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 110 (3c) of IVDR, are shown below:

- 31. December 2027, for devices certified under IVDD
- 31. December 2027, for class D devices;
- 31. December 2028, for class C devices;
- 31. December 2029, for class B devices and for class A devices placed on the market in sterile condition

We reserve the right to invoice any issuance, copies, amendments and / or changes of the confirmation letter according to effort.

For confirmation letter validity see www.tuvsud.com/ps-cert?q=CLI_109172_0005

In case of inquiries please contact medical_devices@tuvsud.com.

On behalf of the Notified Body TÜV SÜD Product Service GmbH,
2025-06-03

TÜV SÜD Product Service GmbH
Medical and Health Services

TÜV SÜD Product Service GmbH
Medical and Health Services

Holly Tang
[Holly Tang \(Jun 3, 2025 15:54 GMT+8\)](#)

Dr. Mostafa Mahmoud
[Dr. Mostafa Mahmoud \(Jun 3, 2025 09:55 GMT+2\)](#)

Holly Tang
Conformity Assessment Responsible (CARE)

Dr. Mostafa Mahmoud
Application Reviewer



Table 1: Devices covered by this letter and for which TÜV SÜD Product Service GmbH is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under IVDR application)	IVDR Device classification (as proposed by the manufacturer and verified during application review)	If the IVDR device is a substitute device, identification of the corresponding IVDD device	IVDD Certificate Reference(s) of the devices under IVDR application, and the NB Identification
SARS-CoV-2 & Influenza A/B & RSV Antigen Combo Test Kit (Colloidal Gold Chromatographic Immunoassay) Basic UDI-DI: 697134876071D29643B	Class C incl. ST/NPT/CDx	N/A	Certification as follows: Certificate No. CeCert/092/W/E.2 Valid until: 26.05.2025 NB: CeCert Sp. Z.o.o NB identification: 2934 TÜV SÜD Product Service GmbH is responsible for appropriate surveillance starting from transfer date: 31.07.2025

Table 2: Devices covered by this letter and for which TÜV SÜD Product Service GmbH is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under IVDR application)	IVDR Device classification (as proposed by the manufacturer and verified during application review)	If the IVDR device is a substitute device, identification of the corresponding IVDD device	IVDD Certificate Reference(s) of the devices under IVDR application, and the NB Identification
Not Applicable			

Confirmation Letter Version History

Date	TÜV SÜD Product Service GmbH internal reference traceable to each version of the letter	Action
2025-05-20	713347729 GZ2443603_CL-S	Initial issue
2025-06-03	713347729 GZ2443603_CL	Transfer to Table 1, Revision 01